

Principle and Methodology

Hellabio Immunofixation Electrophoresis (IFE) kits are intended for *in vitro* diagnosis of monoclonal paraproteins in human serum and other biological samples.

In some pathological cases (multiple myeloma etc) abnormal monoclonal bands appear in electropherogram. The identification of these monoclonal bands can be done by different immunological techniques such as Immunoelectrophoresis and Immunofixation electrophoresis (IFE).

The IFE technique combines zone electrophoresis with immunoprecipitation and is easier and more practical than other techniques. By IFE proteins of sample are first according to their charge separated by electrophoresis on the agarose film, and then they (as antigen) are reacted and fixed with monospecific polyclonal antisera. The monoclonal paraproteins can be heavy chain of IgG, IgA, IgM, (IgD, IgE) and or light chain of kappa or lambda.

Hellabio antisera kit contains Protein fixation solution, the anti heavy chain of IgG, IgA, IgM, and free+bound anti light kappa and lambda chain.

If specific antigen (heavy and or light chain globulins) is present, a characteristic immunoprecipitin band will be formed, which is visualized by staining of the gel with Amido Black.

Required Reagents and Equipment's included in each kit:
[Warning: All reagents from each kit must be used together]

Product	KIT IFE01 / 10 TESTS	KIT IFED01 / 20 TESTS
Agarose Gels	10	10
Electrophoresis Buffer [50X concentrated]	20 ml	20 ml
Staining Solution [5X concentrated]	60 ml	60 ml
Diluent solution [ready to use]	12 ml	12 ml
Destaining solution [500X concentrated]	8 ml	8 ml
Gel Blotter Strips	20	20
Sample Templates	10	10
Antisera templates	10	10
Drying Blotter Sheets	50	50
Gel Blotter Sheets	60	60
Antisera [Anti IgG (γ), IgA (α), IgM (μ), (f+b) κ - and λ -chain]	Goat anti-Human antisera	
Protein fixation solution	1.4 ml	1.4 ml
Instructions for use in English		

All reagents must be used according to the instructions and until the expiration date indicated on the kit

Preparation, storage and stability of the reagents included in the kit:

a) Agarose Gels: Agarose Gels are in non- barbital buffer and other non-reactive ingredients for long stability and optimum resolution of protein fractions. The Gels must be stored at **15 - 25 °C on horizontal position** until the expiration date indicated on the kit. Do not freeze the gels. Carefully discover the gel just before use and follow the instructions of the manual.

b) Diluent Solution: Working electrophoresis buffer + Bromphenol blue and other non-reactive ingredients. Store it at 15 - 25 °C until the expiration date indicated on the kit. **Ready to use.**

c) Electrophoresis Buffer: Non-barbital buffer and other non-reactive ingredients. It is in concentrated solution. It must be stored at 15 - 25 °C until the expiration date indicated on the kit. To prepare working solution dilute the content of the bottle according to the instructions on the bottle. The diluted solution is enough for electrophoresis of all gels of the kit. **The buffer solution is for one use only.** Store the diluted solution at room temperature for 2 months. If crystals appear, place the vial in warm water to dissolve the crystals.

d) Staining Solution: Concentrated Amido Black solution. Store the concentrated solution at 15 - 25 °C until the expiration date indicated on the kit. To prepare working solution dilute the content of the bottle according to the instructions on the bottle. The diluted solution is enough for the staining of all gels of the kit. It should be stored in a closed flask at room temperature until 3 months.

e) Gel Blotter strips: Thin filter paper strips to blot the gel in the application area. Avoid humidity. Blot just for 5 seconds. Avoid humidity.

f) Gel Blotter Sheets: Thin filter paper sheet to blot the entire surface of the gel.

g) Drying Blotter Sheets: Thick filter paper sheet which absorbs the excess humidity of the gel.

h) Destaining solution: 2% citric acid solution. To prepare working solution dilute the content of the bottle according to the instructions on the bottle. Store at room temperature until the expiration date indicated on the kit. For 4L ready to use destaining solution.

i) Washing solution: 0.9% Saline (NaCl).

j) Antisera kit: Each kit contains protein fixation solution, goat anti-Human heavy chain of IgG, IgA and IgM and free+bound anti light chain of kappa (κ) and lambda (λ). Store at -18 °C (long term condition) until the expiration date indicated on the kit.

Additional Reagents and Equipments required and which can be provided by Hellabio:

Power supply, Electrophoresis tank, Staining-destaining baths, HellabioScan (Gel Analyzer).

Limitation / Caution:

- Do not use the Agarose gel if it seems to be dried or infected by fungi.
- Do not freeze the agarose gel.
- Store the kit in horizontal position.
- Do not use hemolytic sera or plasma.

Collection and handling of specimens:

The samples should be collected according to the standard hospital methods.

a) Blood: Analysis is preferably performed on sera in order to avoid the fibrinogen band (in plasma), which migrates in the gamma zone and could lead to a false interpretation. Venous blood is drawn into tube. After centrifugation the supernatant fluid (serum) can be stored at 4-8°C for 48-72 hours. Storage at -20°C may disturb some of the electrophoretic patterns. Care should be taken to prevent haemolysis in the serum because it will cause false elevation in the α_2 and β -fractions. Samples with much precipitate that will not dissolve on warming to 37°C should be either centrifuged or allowed to settle before the clear supernatant fluid is applied to the gel surface.

b) Cerebrospinal fluid: Normally, some 80% of the proteins of cerebrospinal fluid originate from plasma, and the rest have been synthesised locally. The filtration resistance for plasma protein rapidly increases with molecular size and therefore the ratio of large proteins normally is much lower in CSF than in plasma. The protein content of CSF is in the range of 15-40 mg/dl. Therefore it must be concentrated at least 150-fold before electrophoresis.

c) Urine: Analysis of the urinary proteins shows that normal urine contains more kind of proteins which are much diluted and therefore it must be concentrated. For the detection of Bence-Jones proteins, free Kappa and Lambda light chains, the sample has to be concentrated to ≥ 100 mg/dl of total protein and to about 80 mg/dl of total proteins for immunoglobulin

Procedure of Immunofixation electrophoresis

Extreme care must be taken in choosing the appropriate dilution of the serum sample (100 – 200 mg/dl). The incorrect choice of serum dilution may result in either inability to detect a minor monoclonal protein or a prozone effect. When the protein concentration is low (cerebrospinal fluid, urine), it must be concentrated to get a protein concentration at least 100 mg/dl.

- I. Prepare (with proteins diluent solution) freshly dilution of serum sample, so that the concentration of each globulin in corresponded dilution is near 100-200 mg %. For example: when in a case the concentration of IgG globulin's is 3000 mg %, then the sample should be diluted 1:16 for IgG (=187mg%). In case of sample with globulin concentration near the normal level, or with unknown concentration, dilute the sample 1:6(1 vol. serum +5 vol. protein diluent) for PE position and 1:9 for all other positions (1 volume serum +8 volume diluent).
- II. Fill the electrophoresis chamber with adequate volume (it depends on the chamber volume) of electrophoresis buffer.
- III. Take the agarose gel out of its packaging, uncover it from the plastic plate and put it on the backside of the plate in horizontal position.
- IV. Blot the gel for 5" with a gel blotter strip in the zone of sample - application,
- V. Place the sample template on the application zone carefully. Rub the template with forefinger gently to eliminate trapped air bubbles.
- VI. Using a 5-µl pipette, apply 5 µl of serum dilution across each corresponding slit [The application of the samples should be done as quickly as possible. The application slits should not be allowed to dry].
- VII. Let the samples absorb into the gel for 2 minutes and then blot the superfluous sample with a gel blotter strip.
- VIII. Remove both the sample template gently and gel blotter strip and discard them.
- IX. Place the gel into the electrophoresis chamber with the samples on the cathodic side, and run electrophoresis for 20 minutes in 100 volts (the time depends on the kind of the power supply).
- X. After electrophoresis switch off the power supply and place the gel film in a strictly horizontal position.
- XI. Blot the gel surface with a gel blotter sheet and apply on the gel surface the antiserum template. Rub the template with forefinger gently to eliminate air bubbles (very important).
- XII. Apply into the corresponding troughs of the antisera template:

Position	Reagents	USE VOLUME FOR	
		Kit IFE01	Kit IFED01
PE	Fixation solution	90µl	70µl
IgG,IgA,IgM, kappa,lambda	Corresponding antisera	50µl	40µl

- XIII. Incubate the gel film for 10 minutes at room temperature in a moist chamber in strictly horizontal position.
- XIV. Remove the antisera template and discard it.
- XV. Put on the gel one gel blotter sheet and one drying blotter sheet; place a development weight (about 2 kg) for 2 minutes.
- XVI. Soak the gel in saline solution for 5 minutes.
- XVII. Repeat step (XV) and (XVI) three more times.
- XVIII. Clean with a soft paper the back site of the film and dry the gel with hot air (less than 85°C), and stain it for 3 minutes with protein staining solution.
- XIX. Decolonise the gel for 5 minutes in three baths of destaining solutions, subsequently.
- XX. Dry again the gel and evaluate visually the results (see conclusion / troubleshooting).

Results:

The identification and the determination of the monoclonal band should be done by comparing the location of the band(s) with the band(s) with the same location in the PE position (see conclusion / troubleshooting).

Expected values and Evaluation of Results:

In case of monoclonal paraproteinemia monoclonal band(s) will be appear with one or more anti heavy and /or light chain globulin. If you get monoclonal band(s) in PE position, the conclusion / troubleshooting should be as follow:

PE	Anti IgG	Anti IgA	Anti IgM	Anti κ	Anti λ	Conclusion/ Troubleshooting
		-	-	-	-	Monoclonal IgG heavy chain/-
		-	-		-	Monoclonal IgG κ- chain/-
	-	-	-		-	Monoclonal light κ- chain/ Think about IgD and IgE too.
		-	-		-	Monoclonal IgG λ-chain
	-		-		-	Monoclonal IgA κ- chain
	-		-	-	-	Monoclonal IgA heavy chain
	-		-	-		Monoclonal IgA λ- chain
			-	-		Biclonal IgG, IgA IgA λ- chain
	-	-		-	-	Monoclonal IgM heavy chain
	-	-		-		Monoclonal IgM λ- chain
	-	-	-	-	-	Fibrinogen or Heavy chain of IgD or IgE. Do not use plasma for electrophoresis! Examine the case with Anti IgD and IgE antisera too.
						If cryoglobulins, rheumatoid factor, or immuno complexes are present in sample monoclonal bands appear with more than one or with all antisera.

Sensitivity: The sensitivity of fixation solution (on PE position) is 10X higher than with Amido Black. The antisera can detect the specific antigen in a concentration of 80-400mg%.

Specificity: The antisera are monospecific and they do not react with fibrinogen or other human proteins.

REFERENCES:

- Brigden M.L.: Lab Medical International July/August 2000. Merlini G. et al: Biotech Lab International May-June/98.
 Dimitriadis F. et al : Hema 2003 (under press).
 Farhangi, M. and Merlini G.: Seminars in Oncology, 13: 366-79 (1986).
 Kohn J.: J Clin Path, 28: 77-82 (1975).
 Merlini, G.: The Cancer J,8: 173-80 (1995).
 Ritchie, R.Fand Smith, R.: Clin Chem 22,a) 497-99, b) 1735-37 and c) 1982-85 (1976).
 Sun T et al: Amer J Clin Path 72: 5-11 (1979).
 Whicher J.T.et al: Ann Clin Biochem 24: 119-32 (Review).



DIMITRIADIS IOANNIS AND SON PC
 Production of electrophoresis products
 Business Incubator Thermi
 Steliou Kazantzidi 47, 57001, Thermi, Greece
 E-mail: hellabio@hellabio.com
 Phone: + 30 2311 999911
 Fax: +30 2311 999912
 Website:www.hellabio.com